

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. UPHC, East Thambram	<b>Details About Robust Analysis</b>									
<b>Lab No</b>	12259	<b>Detail About Monthly Summary</b>									
<b>Month</b>	May	<b>Detail about SDI</b>									
<b>Year</b>	2023										
<b>Constituent Group</b>	Chemistry III										
		Click on the analyte to view Graphical Data				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 20/05/2023									
		Date of Report Published : 05/06/2023									
SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	172 mg/dL	-0.95	0.44	
2	UREA III	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation )	46	29.96	11.88	3.56	29 mg/dL	-0.27	1.05	
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	68	1.75	7.05	0.12	1.7 mg/dL	-0.41	0.03	
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	119 mg/dL	0.76	0.27	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>