



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult\_new.aspx?resDetId=26.

Lab Name

Govt. Block PHC, Parandur

Constituent Group

Chemistry III

Lab No

7070

PT Item

Lyophilized human serum based

Date of Result Entered :

19/08/2023

Date of Report Published :

01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	197.1 mg/dL	2.39	0.42
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	174	28.29	15.20	4.30	30.4 mg/dL	0.49	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.55 mg/dL	-1.76	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	126.3 mg/dL	2.70	0.22
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**

Govt. Block PHC, Parandur  
The Block Medical Officer Parandur, Parandur Post, Karapettai via, Parandur, Kancheepuram District  
Parandur  
TAMILNADU631552

**MEDICAL OFFICER,**

Govt. Primary Health Centre  
PARANDUR - 631 552  
Kancheepuram District

*Handwritten signature*

*Handwritten signature: Pamela Christudoss*

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in

Dr. Pamela Christudoss  
CMC EQAS Coordinator