



Lab Name Govt. PHC, Thandampalayam

Lab No

12213

Constituent Group Chemistry III

Date of Result Entered :

20/07/2023

PT item Lyophilized human serum based

Date of Report Published :

07/08/2023

| Sl. No. | Analyte | Method | Reference Range | Analyzer Name | No. of Participants | DV | CV | CV | CV | CV | CV | CV | CV | CV |
|---------------------------------|----------------|---------------------------------------|-----------------|--|--|--------|------|------|-------|-------|------|----|----|----|
| 1 | GLUCOSE III | GOD-POD III | | Robonik (Semi automatic / Fully Automatic) | 1876 | 109.35 | 5.55 | 6.07 | 104 | -0.88 | 0.28 | | | |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | | Robonik (Semi automatic / Fully Automatic) | 1850 | 5.83 | 6.00 | 0.35 | 5.99 | 0.46 | 0.02 | | | |
| 3 | CHOLESTROL III | CHOD-PAP III | | Robonik (Semi automatic / Fully Automatic) | 1776 | 113.00 | 4.12 | 4.66 | 104.1 | -1.91 | 0.22 | | | |
| SDI Range Interpretation | | | | | | | | | | | | | | |
| Within -1.00 to +1.00 | | | | | Excellent. | | | | | | | | | |
| Within ±1.01 to ±2.00 | | | | | Good. | | | | | | | | | |
| Within ±2.01 to ±2.99 | | | | | Accept with caution. Warning Signal. | | | | | | | | | |
| Beyond ±3.0 | | | | | Unacceptable performance. Action Signal. | | | | | | | | | |

LAB ADDRESS :
 Govt. PHC, Thandampalayam
 THE MEDICAL OFFICER, KODUMUDI
 Thandampalayam
 TAMILNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

D.K. Srinivas

மருத்துவ அலுவலர்
சென்னை மருத்துவ கல்லூரி
சென்னை

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore