



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC:1023

Lab Name: Govt. UGPHC, Sivagiri  
 Constituent Group: Chemistry III  
 PT Item: Lyophilized human serum based

Lab No: 7382  
 Date of Result Entered: 21/08/2023  
 Date of Report Published: 01/09/2023

Sl. No.	Test Name	Method	Result	Reference Range	Interpretation
1	GLUCOSE III	GOD-POD III	1841	170 mg/dL	-0.62
2	CREATININE III	Jaffes Kinetik - Alkaline Picrate III	1852	1.85 mg/dL	0.76
3	CHOLESTROL III	CHOD-PAP III	1764	109.85 mg/dL	-0.85
SDI Range					
			1841	175.58	5.13
			1852	1.76	6.77
			1764	113.80	4.07
				4.84	0.12
				109.85	0.01
					0.22
Interpretation					
					Excellent.
					Good.
					Accept with caution. Warning Signal.
					Unacceptable performance. Action Signal.

LAB ADDRESS :  
 Govt. UGPHC, Sivagiri  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, ERODE DISTRICT,  
 Sivagiri  
 TAMILNADU638109

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

*Unf*  
 GOVT. UGPHC, SIVAGIRI  
 CLINICAL CHEMISTRY CENTRE  
 SIVAGIRI - 636 105, ERODE (DT).

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*