



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult_new.aspx?resDetId=

Lab Name Govt. Block PHC, Parandur

Constituent Group Chemistry III

PT Item Lyophilized human serum based

Lab No

7070

Date of Result Entered :

20/05/2023

Date of Report Published :

05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	185 mg/dL	0.44	0.44
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	163	27.91	13.39	3.74	28.9 mg/dL	0.26	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.84 mg/dL	0.61	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	116.2 mg/dL	0.27	0.27
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. Block PHC, Parandur
"The Block Medical Officer Parandur, Parandur Post, Karapettai via, Parandur, Kancheepuram District
Parandur
TAMILNADU631552

MEDICAL OFFICER,
Govt. Primary Health Centre
PARANDUR - 631 552
Kanchipuram District

LB

Pamela Christudoss

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in

Dr. Pamela Christudoss
CMC EQAS Coordinator