



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. PHC, Ramapattnam

Lab No

12044

Constituent Group Chemistry III

Date of Result Entered :

19/07/2023

PT Item Lyophilized human serum based

Date of Report Published :

07/08/2023

S.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	104 mg/dL	-0.88	0.28
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	184	51.41	11.14	5.73	43.2 mg/dL	-1.43	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.12 mg/dL	0.83	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	113 mg/dL	0.00	0.22
<b>SDI Range Interpretation</b>										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

### LAB ADDRESS :

Govt. PHC, Ramapattnam  
THE MEDICAL OFFICER, POLLACHI (NORTH)  
Coimbatore  
TAMILNADU

### Coordinator Contact Details:

Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christodoss*

Dr. Pamela Christodoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore