



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. PHC, Ramapattinam

Lab No 12044

Constituent Group Chemistry III

Date of Result Entered : 21/06/2023

PT Item Lyophilized human serum based

Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SD	%
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	305 mg/dL	1.24	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	192	32.69	12.96	4.24	32 mg/dL	-0.16	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	3.2 mg/dL	3.68	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	108 mg/dL	-1.13	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Ramapattinam
THE MEDICAL OFFICER, POLLACHI (NORTH)
Coimbatore
TAMILNADU

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Parvula Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore