



## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12140	MEDICAL DEVICES	Page No:1 of 1
Issue Date	30-09-2023	
SRF No & Date	SRF/23/00468-0006 & 29-09-2023	
Receipt Date	29-09-2023	
Calibration Date	29-09-2023	
Calibration Due	28-09-2024	

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
POONDI BAZAR,EAST TAMBARAM,600059.

**Details of Device Under Calibration (DUC)**

Description : SEMI AUTO ANALYZER	Make & Model : ROBONIK & PRIETEST TOUCH	Range : MULTI	Sr. No : ATCD0471220RBK
Resolution : MULTI	Identification No : NA	DUC Condition : SATISFACTORY	Location : LABORATORY

**Environmental Conditions & Calibration Procedure Details**

Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

**ELECTRICAL SAFETY**

**RESULTS**

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	<b>Insulation Resistance</b> >20MΩ	<b>Measured values in MΩ</b> 87	<b>Uncertainty in % ( ± )</b> 13.92
2	<b>Earth Leakage</b> <5000µAfor B,BF,CF	<b>Measured values in µA</b> 195	<b>Uncertainty in % ( ± )</b> 9.3
3	<b>Enclosure Leakage</b> <500µAfor B,BF,CF	<b>Measured values in µA</b> 241	<b>Uncertainty in % ( ± )</b> 8.4

**REMARKS**

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

*M. Adhiban*  
(Calibration Engineer)  
M.ADHIBAN

*C. Shanmugaraaj*  
Technical Manager  
(C.SHANMUGARAJ)

Chief Executive

