




CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name Govt. PHC, Ramapattnam

Lab No

12044

Constituent Group Chemistry III

Date of Result Entered :

20/04/2023

PT Item Lyophilized human serum based

Date of Report Published :

06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	79 mg/dL	-5.61	0.34
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	169	42.32	10.68	4.52	25 mg/dL	-3.83	0.70
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.2 mg/dL	0.23	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	113 mg/dL	0.24	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ± 1.01 to ± 2.00				Good.						
Within ± 2.01 to ± 2.99				Accept with caution. Warning Signal.						
Beyond ± 3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Ramapattnam
THE MEDICAL OFFICER, POLLACHI (NORTH)
Coimbatore
TAMILNADU

Coordinator Contact Details:
Email: clnqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore