



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Alampatti
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 12436
Date of Result Entered: 18/07/2023
Date of Report Published: 07/08/2023

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOO-POD III	Robonik (Semi automatic / Fully Automatic)	1876	106.35	5.65	9.97	116.9 mg/dL	1.67	0.26
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1886	8.53	6.06	0.31	8.4 mg/dL	-1.23	0.69
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.09	6.12	4.64	103 mg/dL	-1.07	0.22
SDI Range										
Within -1.00 to +1.00										
Excellent										
Within ±1.01 to ±2.00										
Good										
Within ±2.01 to ±3.99										
Accept with caution. Warning Signal.										
Beyond ±3.0										
Unacceptable performance. Action Signal.										
Interpretation										

LAB ADDRESS :
Govt. PHC, Alampatti
CLINICAL LABORATORY
Alampatti
TAMILNADU

Pamela Christudas
Dr. Pamela Christudas
CMC EQAS Coordinator
Christian Medical College, Vellore

Co-ordinator Contact Details:
Email: clinlq@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****