



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name Govt. Block PHC, Akkur
 Constituent Group Chemistry III
 PT Item Lyophilized human serum based

Lab No 7528
 Date of Result Entered: 19/06/2023
 Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SDI			
1	GLUCOSE III	GOD-POD III	Biosystems	36	280.91	4.87	11.42	289 mg/dL	0.74	3.81
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	39	2.67	6.98	9.17	2.7 mg/dL	0.17	0.96
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	113.79	4.23	4.81	118 mg/dL	0.88	1.98

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Akkur

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT.

Akkur

TAMILNADU631701

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

V.K. Murugesan
 SST SURGEON
 J.G.PHC AKKUR
 PIN - 631 701

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****