

Lab No: 10874 [Back to Search Page](#)

Month: April

Year: 2023

Constituent Group: Chemistry III

Date of Result Entered: 19/04/2023

Date of Report Published: 06/05/2023

Sl.No	Constituent Name	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	75	116.67	7.36	8.67			2.93
			Transasia / Erba	19	118.42	8.46	10.04			4.81
			Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.84	7.21	122 mg/dL	0.35	0.34
			Any Analyser (Automation / Semi Automation)	74	118.33	5.77	6.85			1.58
2	CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Biosystems	34	2.16	6.66	0.15			0.96
			Transasia / Erba	19	2.14	6.03	0.13			0.96
			Robonik (Semi automatic / Fully Automatic)	1829	2.16	6.09	0.15	2.2 mg/dL	0.23	0.01
			Any Analyser (Automation / Semi Automation)	85	2.14	7.02	0.15			0.94
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03			2.87
			Transasia / Erba	20	111.15	4.25	4.72			2.11
			Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	113 mg/dL	0.24	0.28
			Any Analyser (Automation / Semi Automation)	143	111.32	4.31	4.79			0.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within +1.01 to +2.00	Good
Within +2.01 to +2.99	Accept with caution / Warning Signal