



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name Govt. Block PHC, Akkur
Constituent Group Chemistry III
PT item Lyophilized human serum based

Lab No 7528
Date of Result Entered : 19/08/2023
Date of Report Published : 01/09/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|---------------|--------------------|--------|--------------|------|------------|------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Biosystems | 34 | 172.03 | 5.51 | 9.48 | 176 mg/dL | 0.42 | 3.25 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Biosystems | 35 | 1.76 | 6.47 | 0.11 | 1.8 mg/dL | 0.35 | 0.04 |
| 3 | CHOLESTROL III | CHOD-PAP III | Biosystems | 33 | 114.98 | 3.90 | 4.49 | 118 mg/dL | 0.67 | 1.56 |

| SDI Range | Interpretation |
|---------------------------------|------------------------------------------|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :
Govt. Block PHC, Akkur
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT.
Akkur
TAMILNADU 631701

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Vk M...
SS1 SURGEON
U.G.PHC AKKUR
PIN - 631 781

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****