

# CHRISTIAN MEDICAL COLLEGE

External Quality Assurance Scheme - First Monthly Summary

## CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. PHC, Alampatti  
City: Alampatti  
Group: Chemistry III

Lab No: 52456

Date of Result Entered: 16/08/2023

Date of Report Published: 01/09/2023

Lyophilized human serum based

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	SD	CV	SD	CV	SD
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	5.13	0.06	5.13	0.06	5.13	0.06
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	6.77	0.12	6.77	0.12	6.77	0.12
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	4.07	0.04	4.07	0.04	4.07	0.04
SDI Range										
Within -1.00 to +1.00										
Within ±1.01 to ±2.00										
Within ±2.01 to ±2.99										
Beyond ±3.0										
Interpretation										
Excellent.										
Good.										
Accept with caution. Warning Signal.										
Unacceptable performance. Action Signal.										

LAB ADDRESS :  
Govt. PHC, Alampatti  
CLINICAL LABORATORY  
Alampatti  
TAMILNADU

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christodoss*  
Dr. Pamela Christodoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*