

### CALIBRATION REPORT


Report No : HBMS/CAL/23-24		Calibration Date : 11.09.2023
		Calibration Valid: 10.03.2024
<b>I.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, GOVT. PRIMARY HEALTH CENTRE, BELUKURICHI, SENDAMANGALAM MAIN ROAD, NAMAKKAL-637402.
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOTECH
ii	Model	BDI-152
iii	Serial No	20111131
iv	Location	LABORATORY

### CALIBRATION WORKSHEET

#### Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1001	1
2	1200	1201	1
3	1500	1502	2
4	2000	1999	-1

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range



S.No	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	95.0	95.1	0.1
2	18.0	18.2	0.2
3	15.0	15.1	0.1
4	28.0	28.2	0.2

Acceptance deviation:  $\pm 0.2$  seconds

### Test equipment used

The calibration was performed by using Digital Multimeter & PT - 100 sensor, Digital Timer (Stop Watch) and Digital Tachometer ( non - contact ) measured has been inspected in accordance with our inspection procedures to meet manufacturer's specification.

### Remarks

This certificate may not be reproduced other than in full except prior permission of laboratory.

Verified by :

