



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name Govt. UPHC, Selvapuram Lab No 10889
Constituent Group Chemistry III Date of Result Entered : 19/08/2023
PT item Lyophilized human serum based Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	23	173.17	4.83	8.36	174 mg/dL	0.10	3.49
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	21	1.74	6.34	0.11	1.9 mg/dL	1.45	0.05
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	17	111.42	3.18	3.54	126 mg/dL	4.12	1.72

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Selvapuram
THE MEDICAL OFFICER, D.NO-19 PERUR MAIN ROAD, SELVAPURAM
Coimbatore
TAMILNADU641026

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

Dr. V. K. R.
Women Health Officer
Selvapuram Urban Health Centre
Coimbatore - 641 006