

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

m Lab No 10532
Date of Result Entered : 20/08/2023
um based Date of Report Published : 01/09/2023

Sample Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
				CV	SD			
	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	177 mg/dL	0.16	0.42
Salicylate	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.74 mg/dL	-0.17	0.01
	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	107 mg/dL	-1.47	0.22

Interpretation
Excellent.
Good.
Accept with caution. Warning Signal.
Unacceptable performance. Action Signal.

Details:
cmcre.ac.in
Phone: +91-420-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

[Signature]
20/9/23

ASSISTANT SURGEON
Govt. Primary Health Centre
Sembakkam-603108.