



CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/12441	MECHANICAL	Page No:1 of 1	
Issue Date	30-09-2023			
SRF No & Date	SRF/23/00535-0001 & 30-09-2023			
Receipt Date	30-09-2023			
Calibration Date	30-09-2023			
Calibration Due	29-09-2024			
Customer Name & Address				
GOVERNMENT URBAN PRIMARY HEALTH CENTRE, NESAVALOR-641602, TIRUPPUR DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	: CENTRIFUGE	Make & Model	: REMI & C-854/4	
Range	: 3500 RPM	Sr. No	: ZBHN-22429	
Resolution	: 1 RPM	Identification No	: NA	
DUC Condition	: Satisfactory	Location	: LABORATORY	
Environmental Conditions & Standard Operating Procedure Details				
Environmental Details	Temperature: 24.3 °C	Relative Humidity	54 % Rh	
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	Lutron & DT-2238	JRPM-CCTR-A&S-2022-0036	03-10-2023

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)
	RPM	RPM	RPM	%
1	1000	999.9	0.1	4.2
2	2000	1998	2	4.2
3	3000	2999	1	4.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



Authorised Signatory

C. Sivabalan
Quality Manager
(C.SIVABALAN) Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12442	MEDICAL DEVICES	Page No:1 of 1
Issue Date	30-09-2023	
SRF No & Date	SRF/23/00535-0002 & 30-09-2023	
Receipt Date	30-09-2023	
Calibration Date	30-09-2023	
Calibration Due	29-09-2024	

Customer Name & Address

GOVERNMENT URBAN PRIMARY HEALTH CENTRE,
NESAVALOR-641602,TIRUPPUR DISTRICT.

Details of Device Under Calibration (DUC)

Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH
Range	MULTI	Sr. No	ATCD1320818RBK
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	97	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)
	<5000μAfor B,BF,CF	125	8.0
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500μAfor B,BF,CF	235	7.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



C.S.
Technical Manager
C.SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12443	MEDICAL DEVICES	Page No:1 of 1
Issue Date	30-09-2023	
SRF No & Date	SRF/23/00535-0003 & 30-09-2023	
Receipt Date	30-09-2023	
Calibration Date	30-09-2023	
Calibration Due	29-09-2024	

Customer Name & Address
 GOVERNMENT URBAN PRIMARY HEALTH CENTRE,
 NESAVALOR-641602, TIRUPPUR DISTRICT.

Details of Device Under Calibration (DUC)			
Description	ELECTRICAL SAFETY (CELL COUNTER)	Make & Model	SYSMEX & XP-300
Range	MULTI	Sr. No	B8015
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS			
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ 97	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 125	Uncertainty in % (±) 12.3
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 229	Uncertainty in % (±) 8.6

- REMARKS**
- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 - The user should determine the suitability of the instrument for its intended use.
 - The recalibration interval should be determined on the user requirement.
 - The results stated in this certificate relate only to the item calibrated.
 - The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
 - Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

M. Adhiban
 (Calibration Engineer)
 M.ADHIBAN



C. Shanmugaraaj
 Technical Manager
 C.SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12444		MEDICAL DEVICES		Page No:1 of 1
Issue Date	30-09-2023			
SRF No & Date	SRF/23/00535-0004 & 30-09-2023			
Receipt Date	30-09-2023			
Calibration Date	30-09-2023			
Calibration Due	29-09-2024			
Customer Name & Address				
GOVERNMENT URBAN PRIMARY HEALTH CENTRE, NESAVALOR-641602, TIRUPPUR DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	ELECTRICAL SAFETY (MICROSCOPE)	Make & Model	KWALITY & KXB-1005	
Range	MULTI	Sr. No	3929	
Resolution	MULTI	Identification No	NA	
DUC Condition	SATISFACTORY	Location	LABORATORY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ 88	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 119	Uncertainty in % (±) 8.1
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 236	Uncertainty in % (±) 7.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



C. Shanmugaraj
Technical Manager
C.SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.