

CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



Sembakkam

Lab No

10532

III

Date of Result Entered :

20/06/2023

d human serum based

Date of Report Published :

04/07/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	L
				CV	SD			
POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	309 mg/dL	1.51	0.1
Kinetic - Alkaline Picrate	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.1
-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	111 mg/dL	-0.54	0.1

Interpretation	
	Excellent.
	Good.
	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

JRAM(DIST)

For Contact Details:
Email: qc@cmcvellore.ac.in
Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****


ASSISTANT SURGEON
Govt. Primary Health Centre
Sembakkam-603108.