



CALIBRATION REPORT

| | | |
|---|-----------------|---|
| Report No : HBMS/CAL/23-24 | | Calibration Date: 28.09.2023 |
| | | Calibration valid: 27.03.2024 |
| 1.1.CUSTOMER DETAILS: | | |
| Name and Address of the organization | | M/s, URBAN PRIMARY HEALTH CENTRE, SELVAPURAM,COIMBATORE. |
| 1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | | |
| EQUIPMENT NAME | | MICRO PIPETTE |
| i | Manufactured by | THERMOSCIENTIFIC |
| ii | Model | FINNPIPETTE F3 |
| iii | Value | 100µl-1000µl |
| iv | Serial No | RW02771 |

Calibration standards used :


| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022 | 06-OCT-2023 |

Calibration test Report.

| Number of measurement | Channel 1 (Volume) | | |
|-----------------------|--------------------|--------|---------|
| | 100µl | 500µl | 1000µl |
| 1 | 100.03 | 500.06 | 1000.14 |
| 2 | 100.04 | 500.09 | 1000.12 |
| 3 | 100.05 | 500.02 | 1000.10 |
| 4 | 100.09 | 500.08 | 1000.08 |
| 5 | 100.00 | 499.94 | 1000.09 |
| 6 | 100.08 | 500.07 | 1000.12 |

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com



| CALIBRATION REPORT | |
|---|---|
| Report No : HBMS/CAL/23-24 | Calibration Date: 28.09.2023 |
| | Calibration valid: 27.03.2024 |
| 1.1.CUSTOMER DETAILS: | |
| Name and Address of the organization | M/s, URBAN PRIMARY HEALTH CENTR, SELVAPURAM, COIMBATORE. |
| 1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | |
| EQUIPMENT NAME | MICRO PIPETTE |
| i Manufactured by | ERBA |
| ii Model | ---- |
| iii Value | 100µl-1000µl |
| iv Serial No | 0F508304 |


Calibration standards used :


| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022 | 06-OCT-2023 |

Calibration test Report.

| Number of measurement | Channel 1 (Volume) | | |
|-----------------------|--------------------|--------|---------|
| | 100µl | 500µl | 1000µl |
| 1 | 100.03 | 500.06 | 1000.14 |
| 2 | 100.04 | 500.09 | 1000.12 |
| 3 | 100.05 | 500.02 | 1000.10 |
| 4 | 100.09 | 500.08 | 1000.08 |
| 5 | 100.00 | 499.94 | 1000.09 |
| 6 | 100.08 | 500.07 | 1000.12 |

For Heart Beat Medical Systems,


(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com



CALIBRATION REPORT

| | | |
|---|-----------------|---|
| Report No: HBMS/CAL/23-24 | | Calibration Date: 28.09.2023 |
| | | Calibration Valid: 27.03.2024 |
| I.1. CUSTOMER DETAILS: | | |
| Name and Address of the organization | | M/s, URBAN PRIMARY HEALTH CENTRE, SELVAPURAM COIMBATORE. |
| I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | | |
| EQUIPMENT NAME | | PIPETTE |
| I | Manufactured by | ERBA |
| II | Model | 5µl - 50µl |
| III | Serial No | 0C494875 |
| IV | Location | LABORATORY |

Calibration standards used :


| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022 | 06-OCT-2023 |

Details of the machine: machine calibrated and checked output.

| Number of measurement | Channel I (Volume) | | |
|-----------------------|--------------------|-------|-------|
| | 05µl | 20µl | 50µl |
| 1 | 05.05 | 20.08 | 50.97 |
| 2 | 05.06 | 20.07 | 20.02 |
| 3 | 05.01 | 20.05 | 20.96 |
| 4 | 05.97 | 20.99 | 20.02 |
| 5 | 05.01 | 20.08 | 20.09 |
| 6 | 05.02 | 20.95 | 20.04 |

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com