

**CALIBRATION CERTIFICATE**

Certificate No: SBS/CL/23/12128

Page. No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,  
PACHAPALAYAM-638701, TIRUPPUR DISTRICT.

SRF No.	SRF/23/00485-0002
SRF Date	30-09-2023
Date of Receipt	28-09-2023
Date of Calibration	30-09-2023
Due Date for Calibration	29-09-2024
Issue Date	01-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	MICROLUX
Range	5-50µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	10-50µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.7	Humidity (%RH)	54	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Result of Calibration in µl


Z Factor: 1.00319


SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10	9.91	9.93	9.92	9.94	9.96	9.94	-0.06	0.02	0.47
		9.94	9.95	9.93	9.95	9.95				
2	30	29.96	29.97	29.98	29.99	29.95	29.95	-0.05	0.03	0.47
		29.90	29.91	29.93	29.94	29.95				
3	50	49.85	49.84	49.83	49.85	49.86	49.86	-0.14	0.02	0.47
		49.86	49.87	49.88	49.89	49.90				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

  
(Calibration Engineer)  
M.DINESH



Authorised by:

  
(Quality Manager/Chief Executive)  
C.SIVABALAN

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12127

Page. No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, PACHAPALAYAM-638701, TIRUPPUR DISTRICT.	SRF No. SRF Date Date of Receipt Date of Calibration Due Date for Calibration Issue Date	SRF/23/00485-0001 30-09-2023 28-09-2023 30-09-2023 29-09-2024 01-10-2023
--	---	---

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	MICROLUX
Range	100-1000µl	Model	NA
Resolution	10µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.6	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl


SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.95	99.89	99.87	99.86	99.85	99.87	-0.13	0.03	0.47
		99.86	99.87	99.86	99.87	99.85				
2	500	499.97	499.98	499.95	499.96	499.95	499.95	-0.05	0.02	0.47
		499.92	499.94	499.93	499.94	499.95				
3	1000	999.95	999.97	999.95	999.96	999.97	999.96	-0.04	0.01	0.47
		999.95	999.97	999.96	999.97	999.98				

Remarks

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

  
 (Calibration Engineer)  
 MADHIBAN



Authorised by:

  
 (Quality Manager/Chief Executive)  
 C.SIVABALAN

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12296	MEDICAL DEVICES	Page No:1 of 1
Issue Date	30-09-2023	
SRF No & Date	SRF/23/00485-0003 & 29-09-2023	
Receipt Date	29-09-2023	
Calibration Date	29-09-2023	
Calibration Due	28-09-2024	

**Customer Name & Address**  
 GOVERNMENT PRIMARY HEALTH CENTRE,  
 PACHAPALAYAM-638701, TIRUPPUR DISTRICT.

### Details of Device Under Calibration (DUC)

Description : SEMI AUTO ANALYZER	Make & Model : ROBONIK & PRIETEST TOUCH
Range : MULTI	Sr. No : ATCD3250321RBK
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : BIO-CHEMISTRY

### Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

### Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY

#### RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	<b>Insulation Resistance</b> >20MΩ	<b>Measured values in MΩ</b> 100	<b>Uncertainty in % (±)</b> 13.92
2	<b>Earth Leakage</b> <5000µAfor B,BF,CF	<b>Measured values in µA</b> 196	<b>Uncertainty in % (±)</b> 7.4
3	<b>Enclosure Leakage</b> <500µAfor B,BF,CF	<b>Measured values in µA</b> 201	<b>Uncertainty in % (±)</b> 7.3

#### REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
 (Calibration Engineer)  
 M.RAGUL



  
 Technical Manager  
 C.SHANMUGARAJ

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

## CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/12297	MECHANICAL	Page No:1 of 1	
Issue Date	30-09-2023			
SRF No & Date	SRF/23/00485-0004 & 29-09-2023			
Receipt Date	29-09-2023			
Calibration Date	29-09-2023			
Calibration Due	28-09-2024			
<b>Customer Name &amp; Address</b>				
GOVERNMENT PRIMARY HEALTH CENTRE, PACHAPALAYAM-638701, TIRUPPUR DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description	: CENTRIFUGE	Make & Model	: B.D INSTRUMENTATION & BD-152	
Range	: 3500 RPM	Sr. No	: NA	
Resolution	: 1 RPM	Identification No	: NA	
DUC Condition	: Satisfactory	Location	: LABORATORY	
<b>Environmental Conditions &amp; Standard Operating Procedure Details</b>				
Environmental Details	Temperature: 25.3°C	Relative Humidity	55% Rh	
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE	
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	JRPM-CCTR-A&S-2023-0013	09-06-2024

**CALIBRATION RESULTS**

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	1000	999.2	0.8	4.2
2	2000	1999.6	0.4	4.2
3	3000	2999.7	0.3	4.2

**REMARKS**

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
(Calibration Engineer)  
M.RAGUL



Authorised Signatory

 (Quality Manager)  (Chief Executive)  
(C.SIVABALAN)

### SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.