

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Natham  
 Lab No: 12560  
 Month: August  
 Year: 2023  
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

All Analyser Result  Print

Date of Result Entered : 19/06/2023

Date of Report Published : 01/09/2023

SL.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation )	64	174.34	5.46	9.51	171 mg/dL	4.35	2.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	62	1.76	7.18	8.13	2.2 mg/dL	3.38	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	149	113.96	4.36	4.97	118 mg/dL	0.51	0.51
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						