home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=...

## VIEW LAB MONTHLY SUMMARY

**Lab Name** Govt. UPHC, LRGR

**Lab No** 10874

**Month** March

**Year** 2023

**Constituent Group** Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result | Print

Date of Result Entered : 18/03/2023

Date of Report Published : 0[illegible]/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | EV | Participants | | Your Value | SDI | U |
|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik ( Semi automatic / Fully Automatic ) | 1878 | 91.48 | 7.46 | 6.82 | 102 mg/dL | 1.54 | 0.31 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1851 | 1.33 | 9.08 | 0.12 | 1.3 mg/dL | -0.25 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik ( Semi automatic / Fully Automatic ) | 1761 | 110.93 | 5.23 | 5.80 | 119 mg/dL | 1.39 | 0.28 |

| SDI Range | Interpretation |
|---|---|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |