

## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO: SBS/CL/23/12416</b>		<b>MEDICAL DEVICES</b>		<b>Page No:1 of 1</b>
Issue Date	30-09-2023			
SRF No & Date	SRF/23/00528-0003 & 30-09-2023			
Receipt Date	30-09-2023			
Calibration Date	30-09-2023			
Calibration Due	29-09-2024			
<b>Customer Name &amp; Address</b>				
GOVERNMENT URBAN PRIMARY HEALTH CENTRE, AKKUR-631701,TIRUVANNAMALAI DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description	ELECTRICAL SAFETY (CELL COUNTER)	Make & Model	SYSMEX & 100	
Range	MULTI	Sr. No	NA	
Resolution	MULTI	Identification No	NA	
DUC Condition	SATISFACTORY	Location	LABORATORY	
<b>Environmental Conditions &amp; Calibration Procedure Details</b>				
Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE	
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY

#### RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	<b>Insulation Resisitance</b> >20MΩ	<b>Measured values in MΩ</b> 97	<b>Uncertainty in % ( ± )</b> 13.92
2	<b>Earth Leakage</b> <5000µAfor B,BF,CF	<b>Measured values in µA</b> 125	<b>Uncertainty in % ( ± )</b> 12.3
3	<b>Enclosure Leakage</b> <500µAfor B,BF,CF	<b>Measured values in µA</b> 229	<b>Uncertainty in % ( ± )</b> 8.6

#### REMARKS


- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
(Calibration Engineer)  
E.ESWAR



  
Technical Manager  
C.SHANMUGARAJ

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.