

<u>CA</u>	LIBRAT	TION CER	TIFICA	ITE	
CERTIFICATE NO: SBS/CL/23/03971		MEDICAL DEVICES			Page No:1 of 1
Issue Date		07-04-2023			
SRF No & Date		SRF/23/00146-0002 & 06-04-2023			
Receipt Date		06-04-2023			
Calibration Date		06-04-2023			
Calibration Due	05-04-2024				
Customer Name & Address					
GOVERNMENT PRIMARY HEALTH CEN	TRE,				
PULIYUR-602024,THIRUVALLUR DISTRI	CT.				
	Details of D	evice Under Calibr	ration (DUC)		
Description : SEMI AUTO ANALYZER		Make & Model		PACE -28	
Range :		Sr. No		8A-CL-4639	
Resolution :		Identification No			
DUC Condition : SATISFACTORY		Location		LAB	
Envir	onmental Cond	ditions & Calibratio	on Procedure	Details	
Environmental Details Temperate	ire: 26.4° C	Relative Hu	umidity	54% RH	
Calibration Procedure No SBS/CP/N	D/20	Calibration	done at	ONSITE	
	Refer	ence Standards D	etails		
S.No Description	Make/ SI No	I No: Certif		No	Validity
1 Electrical Safety Analyser Rigel Medic		al & 44L-1059	TSC/22-23	3/7400-3	10-08-202

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1.	Insulation Resistance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	95	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	123	5.9
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)
L.	<500µAfor B,BF,CF	215	11.4

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer)

S.MURALI



		COATS /
Γ	1	Quality Manager
		(D.VETRI SELVI)

ľ
Chief Executive

Authorised Signatory