



CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 11.09.2023 Calibration Valid: 10.03.2024
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVT. PRIMARY HEALTH CENTRE, BELUKURICHI, SENDAMANGALAM MAIN ROAD, NAMAKKAL-637402.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I Manufactured by	THERMOSCIENTIFIC
II Model	-----
III Value	10µl
IV Serial No	GPPHC/BLI/PIP/002

Calibration standards used :


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel I (Volume)
	10µl
1	10.05
2	10.06
3	10.01
4	9.97
5	10.01
6	10.02

For Heart Beat Medical Systems

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com