

		CALIB	RATIO	CERTI	FICATE			
CERTIFICATE NO:	SBS/CL/	SBS/CL/23/03972			IECHANICAL	Pa	age No:1 of 1	
Issue Date			07-	07-04-2023				
SRF No & Date			SRI	SRF/23/00146-0003 & 06-04-2023				
Receipt Date			06-	06-04-2023				
Calibration Date			06-	06-04-2023				
Calibration Due				05-04-2024				
Customer Name & A	ddress				The same			
GOVERNMENT PRIM	MARY HEALTH	CENTRE,				R.T.		
PULIYUR-602024,TH	IRUVALLUR [DISTRICT.						
		Deta	ils of Device U	Inder Calibratio	on (DUC)			
Description :	: CENTRIFUGE			Make & Model : LABOTECH				
Range :	3500	RPM	Sr.	No				
Least Count :	70	RPM	Ide	ntification No				
DUC Condition	Condition : Satisfactory			Location : LAB				
	Envir	onmental Co	onditions & St	andard Operat	ing Procedur	e Details		
Environmental Details Temperature		ure: 25.8 °C	e: 25.8 °C Relative H		50% Rh			
Calibration Procedure No SBS/CP/ML		1L/04	04 Calibration		done at ONSITE			
4	ALCO PARTY OF THE	Washington and the same of the	Reference S	tandards Deta	ils			
S.No Description	lo Description			Make/ SI No:		Certificate No		
1 Digital Tachon	Digital Tachometer		LINE SEIKI	LINE SEIKI / 175-0034V		TMS/22/63 01		

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	1000	999.6	0.4	4.2
2	2000	1999.7	0.3	4.2
3	3500	3499.7	0.3	4.2

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Authorised Signatory		
Quality Manager Chief Execu	tive	
(D.VETRI SELVI)		
	Quality Manager Chief Execu	