



CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 11.09.2023 Calibration Valid: 10.03.2024
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVT. PRIMARY HEALTH CENTRE, BELUKURICHI, SENDAMANGALAM MAIN ROAD, NAMAKKAL-637402.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I Manufactured by	THERMOSCIENTIFIC
II Model	FINNPIPETTE F3
III Value	1000µl
IV Serial No	UPHC/BLI/PIP/003

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	1000µl
1	1000.09
2	1000.06
3	1000.05
4	999.99
5	1000.03
6	1000.01

For Heart Beat Medical Systems,

(Authorized signatory)

