



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12419

Page. No: 1 of 1

Customer Name & Address

GOVERNMENT URBAN PRIMARY HEALTH CENTRE,
AKKUR-631701, TIRUVANNAMALAI DISTRICT.

SRF No.	SRF/23/00528-0006
SRF Date	30-09-2023
Date of Receipt	29-09-2023
Date of Calibration	30-09-2023
Due Date for Calibration	29-09-2024
Issue Date	30-09-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	NA	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.6	Humidity (%RH)	54	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10.0	9.96	9.95	9.98	9.96	9.97	9.97	-0.03	0.01	0.47
		9.98	9.96	9.97	9.98	9.95				
2	50.0	49.86	49.85	49.87	49.86	49.85	49.91	-0.09	0.05	0.47
		49.95	49.96	49.97	49.95	49.95				
3	100.0	99.98	99.97	99.98	99.96	99.98	99.96	-0.04	0.01	0.47
		99.95	99.96	99.94	99.95	99.96				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



Authorised by:

C. Sivabalan
Quality Manager/Chief Executive
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.