

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Samathur Lab No 12046

Constituent Group Chemistry III Date of Result Entered: 19/08/2023

PT item Lyophilized human serum based Date of Report Published: 01/09/2023

OLNI-	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Varia Valua	CDI		
SI.No						CV	SD	Your Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	158.26 mg/dL	-1.92	0.42	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.69 mg/dL	-0.59	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	111.16 mg/dL	-0.57	0.22	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Samathur THE MEDICAL OFFICER,POLLACHI (SOUTH) Samathur TAMILNADU642123

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential