

CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/12445	MECHANICAL	Page No:1 of 1
Issue Date	30-09-2023		
SRF No & Date	SRF/23/00536-0001 & 30-09-2023		
Receipt Date	30-09-2023		
Calibration Date	30-09-2023		
Calibration Due	29-09-2024		

Customer Name & Address

GOVERNMENT URBAN PRIMARY HEALTH CENTRE,
SOOSAIYAPURAM-641601, TIRUPPUR DISTRICT.

Details of Device Under Calibration (DUC)

Description : CENTRIFUGE	Make & Model : REMI & C-854/4
Range : 3500 RPM	Sr. No : ZEFN-16782
Resolution : 1 RPM	Identification No : NA
DUC Condition : Satisfactory	Location : LABORATORY

Environmental Conditions & Standard Operating Procedure Details

Environmental Details	Temperature: 25.3°C	Relative Humidity	55% Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	JRPM-CCTR-A&S-2023-0013	09-06-2024

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	1000	999.4	0.6	4.2
2	2000	1999.7	0.3	4.2
3	3000	2999.6	0.4	4.2

REMARKS



1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By


 (Calibration Engineer)
 M.RAGUL



Authorised Signatory



 (Quality Manager) (Chief Executive)
 (C.SIVABALAN)

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12446	MEDICAL DEVICES	Page No:1 of 1
---------------------------------	-----------------	----------------

Issue Date	30-09-2023
SRF No & Date	SRF/23/00536-0002 & 30-09-2023
Receipt Date	30-09-2023
Calibration Date	30-09-2023
Calibration Due	29-09-2024

Customer Name & Address

GOVERNMENT URBAN PRIMARY HEALTH CENTRE,
SOOSAIYAPURAM-641601, TIRUPPUR DISTRICT.

Details of Device Under Calibration (DUC)

Description : SEMI AUTO ANALYZER	Make & Model : ROBONIK & PRIETEST TOUCH
Range : MULTI	Sr. No : ATCD1360818RBK
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	85	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)
	<5000μAfor B,BF,CF	195	7.4
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500μAfor B,BF,CF	241	7.2

REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory


(Calibration Engineer)
M.RAGUL




Technical Manager
C.SHANMUGARAJ

Chief Executive

CALIBRATION CERTIFICATE

Page No:1 of 1

CERTIFICATE NO: SBS/CL/23/12447	MEDICAL DEVICES	
Issue Date	30-09-2023	
SRF No & Date	SRF/23/00536-0003 & 30-09-2023	
Receipt Date	30-09-2023	
Calibration Date	30-09-2023	
Calibration Due	29-09-2024	

Customer Name & Address

GOVERNMENT URBAN PRIMARY HEALTH CENTRE,
SOOSAIYAPURAM-641601, TIRUPPUR DISTRICT.

Details of Device Under Calibration (DUC)

Description	ELECTRICAL SAFETY (MICROSCOPE)	Make & Model	DEEP VISION & DV-30
Range	MULTI	Sr. No	NA
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 93	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µA for B,BF,CF	Measured values in µA 173	Uncertainty in % (±) 7.5
3	Enclosure Leakage <500µA for B,BF,CF	Measured values in µA 243	Uncertainty in % (±) 7.1

REMARKS


1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory


(Calibration Engineer)
M.RAGUL




 Technical Manager
C.SHANMUGARAJ

Chief Executive