

			CALIB	RATIO	N CE	RTIFI	CATE			
CERTIFICATE NO:		SBS/CL/	SBS/CL/23/03974			THERMAL		Page	No:1 of 1	
Issue I	Date				07-04-2023				ASSOCIATION ROLL ST	
SRF No & Date					SRF/23/00146-0005 & 06-04-2023					
Receipt Date					06-04-2	023				
Calibra	ation Date				06-04-2023					
Calibration Due					05-04-2024					
Custo	mer Name & Add	ress								
GOVE	RNMENT PRIMA	RY HEALTH	CENTRE,							
PULIY	'UR-602024,THIRI	JVALLUR D	DISTRICT.							
			Detai	ls of Device	Under Ca	libration (I	DUC)			
Description :			TEMPERATURE INDICATOR WITH SENSOR OF (REFRIGERATOR)			Make &Model :		GODREJ & RD EDGE 185 CTM4.2		
Range :		2°C to 8°C			Sr. No :			1501230520		
Least	Count :	0.1	°C		Identificati	on No	N. N. T. BANK			
DUC Condition :		Satisfactory		Location :			LAB			
		E	nvironmenta	I Conditions	s & Calibr	ation Proce	edure Details			
Environmental Details			Temperature: 25.6° C		Relative Humidity		52% Rh			
Calibration Procedure No)	SBS/CP/TH/04		Calibration done at		ONSITE			
				Reference	Standard	s Details				
			Make/ SI N	fake/ SI No:		Certificate No		Validity		
Digital Temperature indicator with RTD sensor			Yudian & A	Yudian & Al-500 TD02224192		RI/CERT/0822/727-03		29-09-2023		
CALIE	RATION RESULT	S								

S.No.	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	°C	°C	°C	°C	
1	2.0	1.9	0.4	0.63	
2	6.0	5.8	0.5	0.63	
3	8.0	7.8	0.3	0.63	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 7. Temperature Scale:International Temperature Scale,1990(ITS-90)

(Calibration Engineer) S.MURALI

Calibrated By



Quality Manager (D.VETRI SELVI)

Chief Executive

Authorised Signatory