

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. PHC, Semmedu		<b>Details About Robust Analysis</b>							
<b>Lab No</b>	12051		<b>Detail About Monthly Summary</b>							
<b>Month</b>	August		<b>Detail about SDI</b>							
<b>Year</b>	2023									
<b>Constituent Group</b>	Chemistry III									
Click on the analyte to view Graphical Data			<a href="#">All Analyser Result</a>		<a href="#">Print</a>					
			Date of Result Entered : 20/08/2023							
			Date of Report Published : 01/09/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	160 mg/dL	-1.73	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.9 mg/dL	1.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	108 mg/dL	-1.25	0.22
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

Signature  
30/09/2023.