

			CALI	BRA	TION	CERT	IFICAT	<u>E</u>		
CERTIFICATE NO: SBS/CL/2			3/12441			ME	CHANICAL	Pa	Page No:1 of 1	
Issue Date					30-09-2023					
SRF No & Date					SRF/23/00535-0001 & 30-09-2023					
Receipt Date					30-09-2023					
Calibration Date					30-09-2023					
Calibration Due					29-09-2024					
Custo	mer Name & Ac	dress								
GOVE	RNMENT URBA	N PRIMARY	HEALTH CEN	NTRE,						
VESAV	/ALOR-641602,	TIRUPPUR DI	STRICT.							
			Det	ails of D	evice Ur	der Calibrati	ion (DUC)			
Description : CENTRIFUGE			UGE	Make & Model		:	REMI & C-854/4			
Range : 3500 RPM			RPM Sr. No		Sr. No	) :		ZBHN-22429		
Resolution : 1 RPI			RPM Identi		Identifica	tification No :		NA		
DUC Condition : Satisfactory			ry	Location		ı :		LABORATORY		
		Env	ironmental C	ondition	ns & Star	ndard Operat	ting Procedu	re Details		
Environmental Details Temperature:				e: 24.3 °	: 24.3 °C Relative		umidity 54 % Rh			
Calibration Procedure No SBS/CP				S/CP/ML/04		Calibration done at ONSITE				
				Refer	ence Sta	andards Deta	ils			
S.No	Description			Make/ S	Make/ SI No:		Certificate No		Validity	
1 Digital Tachometer				Lutron & DT-2238		8	JRPM-CCTR-A&S-2022-0036		03-10-2023	
ALIB	RATION RESUL	<u>.TS</u>								
S.No	DEVICE UNDER CALIBRATION READINGS			INSTRUMENTS		DEVIATION		EXPANDED UNCERTAINTY (±)		
	RPM			RPM		RPM		%		
1	100	999.9			0.1		4.2			
2	200	1998		2		4.2				
3 3000			2999		1		4.2			

## REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.

4. The results stated in this certificate relate only to the item calibrated.

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor

k=2.00.

6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

M. Adhibay (Calibration Engineer) M.ADHIBAN



Authorised Signatory

Quality Manager

(C.SIVABALAN)

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.