

			CA	LIBRA	TIO	N CEF	RTIF	-IC	ATE_	
CER	CERTIFICATE NO: SBS/CL/23/12443				MEDICAL DEVICES					Page No:1 of 1
Issue Date				30-09-2023				•		
SRF No & Date				SRF/23/00535-0003 & 30-09-2023						
Receipt Date					30-09-2023					
Calibration Date					30-09-2023					
Calibration Due				29-09-2024						
Cust	omer Name	& Add	iress							
GOV	ERNMENT	URBAN	PRIMARY HEALT	H CENTRE,						
NES	AVALOR-64	1602,T	IRUPPUR DISTRIC	T.						
				Details of I	Device	Under Calib	ration	(DUC	)	
			ELECTRICAL SAF	ETY						
Description : (CELL COUNTER)			Make	& Model		:	SYSMEX & XP-300			
Range : MULTI			Sr. N	0		:	B8015			
Resolution : MULTI			Identi	fication No		:	NA			
DUC	Condition	:	SATISFACTORY		Locat			:	LABORATORY	
			Envi	onmental Con	ditions	& Calibrati	on Pro	cedu	re Details	
Environmental Details Temperature:2						umidity		52% RH		
Calibration Procedure No SBS/CP/MD/29			0/29	9 Calibration		done at ONSIT		ONSITE		
						Standards [	etails			
S.No	Description Make/ SI		Make/ SI No	lo:		Certificate No			Validity	
1	Electrical Safety Analyser Rigel Med		Rigel Medica	cal & 44L-1059		M-230809-16-4			10-08-2024	

		ELECTRICAL SAFETY		
RESULTS	3			
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)	
1	Insulation Resisitance	Measured values in $M\Omega$	Uncertainty in % (±)	
	>20MΩ	97	13.92	
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)	
	<5000µAfor B,BF,CF	125	12.3	
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)	

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.

<500µAfor B,BF,CF

- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards. Calibrated By

Technical Manager C.SHANMUGARAJ

Chief Executive

Authorised Signatory

8.6

(Calibration Engineer) M.ADHIBAN

