

# S J V Scientific Company

20, Railway Lane Street, 1<sup>st</sup> Cross, ChekkanThottam, Ganapathy, Coimbatore – 641 006.

E-mail: [sjv.scientific@gmail.com](mailto:sjv.scientific@gmail.com). Ph no: 97866-90789

GSTIN: 33AJXPR6161K1ZG

Ref 2023/Cal/002/01

To,  
The Medical Officer,  
Govt Primary Health Centre,  
Mathampalayam.

## LABOTECH CENTRIFUGE CALIBRATION:

- Fitted with resiliently mounted motor for vibration free performance.
- A specially wounded 5 step transformer is used for speed regulation.
- The highly sophisticated elegant body of the equipment is made with fiber re-enforced plastic which gives protection against corrosion and electric shock.

### Technical Data:

Maximum Speed :4000 RPM,  
Maximum RCF:1880 x g,  
Motor: 1/20 HP AC/DC universal.  
Serial No : 76399


Comments: The above mentioned product is working as expected.

Recommendation: -NIL-

Calibrated Date : 29.09.2023

Due Date : 28.09.2024

SJV Scientific Company



Sales & Service:

Blood Glucometer - BP Apparatus - Weighing Scale - Nebulizer - ECG Machine - Lab Kits & Chemicals - Surgical Instruments.