



Vanavil Calibrations

Make It Simple, But Significant!

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, ERAIYUR, KALLAKURICHI - 607 201.	Certificate No:	VCPI/CC/23-24/30 004
	Date of Calibration:	29 09 2023
	Date of issue:	30 09 2023
	Date of Receipt:	29 09 2023
	Next cal Date:	28 09 2024

Details of Device Under Calibration			
Description:	MICROPIPETTE	Make:	THERMO SCIENTIFIC
Range:	(100 to 1000) μ l	Model/Type:	FINNPIPETTE F3
Resolution:	0.2 μ l	Condition of DUC:	GOOD
Serial Number:	RW11760	Location:	LABORATORY
ID number:	GPHC/LAB/MP-02	Cal. At:	LAB
Environmental Conditions		Calibration Method Used	
Temperature:	23.5 °C	Calibration Method Used:	ISO 8655 2002 PART-6
Humidity:	49 % Rh	Cal procedure no:	-

Master Details

SL no.	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

RESULT OF CALIBRATION

S NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	100.0	99.89	0.11
2	500.0	499.84	0.16
3	1000.0	999.79	0.21

Remarks:

- 1 DUC is Defined as Device Under Calibration.
- 2 Calibration Frequency decided by the customer.
- 3 The above instrument is working in good condition.

Calibrated By:


C.AKASH

(CALIBRATION ENGINEER)



Authorised by:


M. BALAJI,
Quality Manager

Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Blo Medical Engineers)

S2, Rishi Mahara] Apartment, Indra Nagar, 250 Reddiyar Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranal, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

vanavilcalibrations@gmail.com www.rainbowblomedicalengineers.com | www.vanavilcalibrations.com