

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. Block PHC, Uppathur, Sattur

Lab No

7310

Constituent Group

Chemistry III

Date of Result Entered :

20/07/2023

PT item

Lyophilized human serum based

Date of Report Published :

07/08/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	- 11
31.110						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	104 mg/dL	-0.88	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	128 mg/dL	3.22	0.22

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. Block PHC, Uppathur, Sattur CLINICAL LABORATORY, VIRUDHUNAGAR DISTRICT Uppathur, Sattur TAMILNADU626205

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******