




Trust & Care

# heartbeat

Medical System



### CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date: 23.09.2023
		Calibration Valid: 22.03.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/S. CORPORATION HOSPITAL, UPHC,VELLAKINAR,COIMBATORE.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	MICROPET
ii	Model	FINNEPIPETTE F3
iii	Value	100µl-1000µl
iv	Serial No	UPHC/VK/MP/001

#### Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

#### Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu

CBE - 04, hbmedicalsystem@gmail.com