

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, THIRUVEGAMPET, SIVAGANGAI-630 408		Certificate No: VCPL/CC/23-24/25-005	
		Date of Calibration: 29-09-2023	
		Date of issue: 30-09-2023	
		Date of Receipt: 29-09-2023	
		Next cal Date: 28-09-2024	
Details of Device Under Calibration			
Description:	MICROPIPETTE	Make:	ERIC
Range:	(100 to 1000) μ l	Model/Type:	-
Resolution:	10 μ l	Condition of DUC	GOOD
Serial Number	GPPHC/LAB/PIP/01	Location:	LABORATORY
ID number	-	Cal. At:	LAB
Environmental Conditions		Calibration Method Used	
Temperature	23.2 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	49 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

RESULT OF CALIBRATION

S.NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	100	100.02	-0.02
2	500	499.91	0.09
3	1000	999.87	0.13

Remarks:

1. DUC is Defined as Device Under Calibration.
2. Calibration Frequency decided by the customer.
3. The above instrument is working in good condition.


Calibrated By:


C. AKASH

(CALIBRATION ENGINEER)



Authorised by:


M. BALAJI,
Quality Manager
Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com