



Vanavil Calibrations
Make It Simple, But Significant !

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, THIRUVEGAMPET,SIVAGANGAI-630 408		Certificate No: VCPL/CC/23-24/25-001
		Date of Calibration: 29-09-2023
		Date of issue: 30-09-2023
		Date of Receipt: 29-09-2023
		Next cal Date: 28-09-2024
Details of Device Under Calibration		
Description:	BIOCHEMISTRY ANALYZER	Make: ROBONIK
Range:	-	Model/Type: PRIETEST TOUCH
Resolution:	-	Condition of DUC GOOD
Serial Number	ATCD0440321RBK	Location: LABORATORY
ID number	-	Cal. At: ONSITE
Environmental Conditions		Calibration Method Used
Temperature	26.2 °C	National/International standard: IEC60601-1
Humidity	51 % Rh	Cal procedure no: -

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyzer	6014025	ACCS230581	FLUKE/ESA-612	9-Mar-24

Electrical Safety Report


S.NO	PARAMETERS	MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
1	Earth Leakage	154.6	<5000 μ A For B,BF,CF
		MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
2	Enclosure Leakage	201.2	<5000 μ A For B,BF,CF

Remarks:


- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


C.AKASH
(CALIBRATION ENGINEER)



Authorised by:


M. BALAJI,
Quality Manager
Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyar Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranal, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com