

14/08/23
TEST NAME: GLUCOSE
12-15-07

Alam Pathi - 26

8/28/23, 11:30 AM
8/15/23, 2:17 PM

External Quality Assurance Scheme - View Monthly Summary
External Quality Assurance Scheme - View Monthly Summary

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UGPHC, Mohanur
Lab No: 7454
Month: August
Year: 2023
Constituent Group: Chemistry III

Details About Robust Analysis
Detail About Monthly Summary
Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result Print

Date of Result Entered : 18/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	175 mg/dL	-0.06	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	126 mg/dL	2.63	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

[Signature]
Medical Officer,
Govt. Upgraded Primary Health Centre,
Mohanur - 637 015,
Namakkal District

2023.09.30 18:05