



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name Govt. PHC, Avinasi Lab No 10620
Constituent Group Chemistry III Date of Result Entered : 18/08/2023
PT item Lyophilized human serum based Date of Report Published : 01/09/2023

| SL.No | Analyte | Method / Principle Name | Analyzer Name | No. of Participants | DV | Participants | | Your Value | SDI | SI |
|-------|----------------|---------------------------------------|--|---------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | St | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 172 mg/dL | -0.40 | 0.42 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.82 mg/dL | 0.50 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 115 mg/dL | 0.26 | 0.22 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution, Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. PHC, Avinasi
Attn to : Mrs.LAILA BEEVI CLINICAL LABORATORY , TIRUPUR District .,
Avinasi
TAMILNADU641654

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****