



**CALIBRATION REPORT**

Report No : HBMS/CAL/23-24		Calibration Date : 12.09.2023
		Calibration valid : 11.03.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s, GOVERNMENT PRIMARY HEALTH CENTRE , PETTAI, THIRUNELVELI</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
<b>i</b>	Manufactured by	Remi
<b>ii</b>	Model	R8C
<b>iii</b>	Serial No	ZEAN-03157

**CALIBRATION WORKSHEET**

**LOCATION: LABORATORY**

**Tabulation For – Centrifuge**

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM	Combined uncertainty in RPM in % rdg
1	1000	1000	0	1.12
2	1200	1201	1	1.23
3	1500	1501	1	1.56
4	2000	2005	5	1.175

Acceptable deviation is  $\pm 5$  RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com