

VIEW LAB MONTHLY SUMMARY



CHRISTIAN MEDICAL COLLEGE
 YELLORE
The Semiter of the Platoon Since 1900...

Robust Analysis

Lab Name

Govt. PHC, Avinasi

Lab No

10620

Month

May

Year

2023

Constituent Group

Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

[Click on the analyte to view Graphical Data](#)

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Year Value	SDI	U
						CV	SD	SDI			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	196.8 mg/dL	1.69	0.44	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	2.04 mg/dL	2.12	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	114.16 mg/dL	-0.08	0.27	
SDI Range											
Within -1.00 to +1.00											
Excellent.											
Within ±1.01 to ±2.00											
Good.											
Within ±2.01 to ±2.99											
Accept with caution. Warning Signal.											
Beyond ±3.0											
Unacceptable performance. Action Signal.											

[Links](#)

[Summary Lab Results](#)

[View Lab Monthly Summary](#)

19/05/2023 10:00:00 AM