



# CHRISTIAN MEDICAL COLLEGE VELLORE

In Service of the Nation Since 1900...



## LAB SUMMARY

Govt. PHC, Kovilangulam

12558

June

2023

Sup Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value
					CV	SD	
GLU III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	287 mg/dL
URE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.7 mg/dL
CHOL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	114 mg/dL

CV Range	Interpretation
CV to $\pm 1.00$	Excellent.
CV to $\pm 2.00$	Good.
CV to $\pm 2.99$	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

Site Maintained by  
**Software Solutions**

**MEDICAL OFFICER**  
Govt. Primary Health Centre  
Kovilangulam  
K. Muthi (T.k) Ramanad (D.t)

Best Viewed at