

HEM LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Kovilangulam
Lab No: 12345
Month: July
Year: 2023
Department Group: Chemistry II

[Details About Patient Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analysis to view Graphical Data](#)

[All Analysis Results](#) [Print](#)

Date of Result Entered: 19/07/2023

Date of Report Published: 07/08/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Total Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Substrak / Semi automatic / Fully Automatic	1876	109.35	5.55	6.07	109 mg/dL	-0.04	0.28
2	UREA / URIC II	Jaffee Kinetic / Alkaline Phosph II	Substrak / Semi automatic / Fully Automatic	1858	5.83	6.80	6.35	1.0 mg/dL	-0.09	0.82
3	CHOLESTEROL II	CHOD-PAP II	Substrak / Semi automatic / Fully Automatic	1776	112.80	4.12	8.64	114 mg/dL	0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Labels			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Weekly Summary	Change Password	Home	Sign-out


MEDICAL OFFICER
 Govt. Primary Health Centre
 Kovilangulam,
 Muthi (T.k) Ramanad (D.O)