

## CALIBRATION CERTIFICATE


CERTIFICATE NO:	SBS/CL/23/12303	MECHANICAL	Page No:1 of 1	
Issue Date	30-09-2023			
SRF No & Date	SRF/23/00500-0001 & 30-09-2023			
Receipt Date	30-09-2023			
Calibration Date	30-09-2023			
Calibration Due	29-09-2024			
<b>Customer Name &amp; Address</b>				
GOVERNMENT PRIMARY HEALTH CENTRE, KOVILANGULAM,RAMNAD DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description	CENTRIFUGE	Make & Model	NA & BDI-152	
Range	3500 RPM	Sr. No	NA	
Resolution	1 RPM	Identification No	NA	
DUC Condition	Satisfactory	Location	LABORATORY	
<b>Environmental Conditions &amp; Standard Operating Procedure Details</b>				
Environmental Details	Temperature: 25.5°C	Relative Humidity	52%Rh	
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE	
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	JRPM-CCTR-A&S-2023-0013	09-06-2024

### CALIBRATION RESULTS


S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	500	498.9	1.1	4.2
2	1500	1499.5	0.5	4.2
3	3500	3499.5	0.5	4.2

### REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By  
  
 (Calibration Engineer)  
 P.PRASANNA



Authorised Signatory  
  
 (Quality Manager)     (Chief Executive)  
 (C.SIVABALAN)