



Lab Name: [New Paltz, New York](#)
 Lab No: 12345
 Month: May
 Year: 2023
 Crosscheck Group: Chemistry II

[Details about Subject Analysis](#)
[View Lab Monthly Summary](#)
[Print](#)

[Click on the analysis to view graphical form](#)

[All Analyser Result](#) | [Print](#)
 Date of Result Entered: 20/05/2023

Date of Report Published: 20/05/2023

SL No.	Analyzer	Method / Assay Name	Analyzer Name	No of Participants	CV	Performance CV	Year	SD	N	
1	ULTRA CHEM II	DODD-PDD-II	Autonics : Semi automatic / Fully Autonics : Fully	1000	1.60 %	1.19	2022	1.60 mg/dL	6.78	6.66
2	ULTRA CHEM II	Alkaline Phosphate	Autonics : Semi automatic / Fully Autonics : Fully	1000	1.76	1.49	2022	1.4 mg/dL	-1.21	6.61
3	ULTRA CHEM II	DODD-PAP-II	Autonics : Semi automatic / Fully Autonics : Fully	1000	1.14 %	1.09	2022	1.11 mg/dL	6.58	6.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution: Warning Signal
Beyond ±3.0	Unacceptable performance: Action Signal

City Address	Health Institute	Related Lab Reports	View Lab Monthly Summary
Comments	Storage Personnel	None	Signed

Maintained by
Swiftware Solutions

Last Viewed at: 12:14 A.M.

MEDICAL OFFICER
 Govt. Primary Health Centre
 Kovilangulam,
 Sathyi (T.k) Ramanad (D.L)