

**CALIBRATION REPORT**

Report No : HBMS/CAL/23-24	Calibration Date : 23.09.2023
	Calibration valid : 22.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s. GOVERNMENT PRIMARY HEALTH CENTRE, THALIYUR, COIMBATORE.

1.2. DESCRIPTION OF DEVICE UNDER TEST (DU/T):

EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	I-PETTE
ii Model	
iii Value	10 µl -100 µl
iv Serial No	PHC/TH/MP/001

Calibration standards used :


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)			
	10µl	50µl	100µl	
1	10.03	50.06	100.14	
2	10.04	50.09	100.12	
3	10.05	50.02	100.10	
4	10.09	50.08	100.08	
5	10.00	49.94	100.09	
6	10.08	50.07	100.12	

For Heart Beat Medical Systems.

(Authorized signatory)



Heartbeat Medical Systems
District: Palakkad, Taluk: Changanassery
Village: Changanassery - 641 017
GATEWAY TO CARE

MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com